



Year: 2012

Image quality and radiation dose comparison of prospectively triggered low-dose CCTA: 128-slice dual-source high-pitch spiral versus 64-slice single-source sequential acquisition

Ghadri, J R ; Küest, S M ; Goetti, R ; Fiechter, M ; Pazhenkottil, A P ; Nkoulou, R N ; Kuhn, F P ; Pietsch, C ; von Schulthess, P ; Gaemperli, O ; Templin, C ; Kaufmann, P A

Abstract: Currently 64-multislice computed tomography (MSCT) scanners are the most widely used devices allowing low radiation dose coronary CT angiography (CCTA) with prospective ECG triggering. Latest 128-slice dual-source CT (DSCT) scanners offer prospective high-pitch spiral acquisition covering the heart during one single beat. We compared radiation dose and image quality from prospective 64-MSCT versus high-pitch spiral 128-slice DSCT scanning, as such data is lacking. CCTA of 50 consecutive patients undergoing 128-DSCT ($2 \times 64 \times 0.6$ mm collimation, 0.28 s rotation time, 3.4 pitch, 100-120 kV tube voltage and 320 mAs tube current-time product) were compared to CCTA of 50 heart rate (HR) and BMI matched patients undergoing 64-MSCT (64×0.625 mm collimation, 0.35 s rotation time, 100-120 kV tube voltage and 400-650 mA tube current). Image quality was rated on a 4-point scale by two independent cardiac imaging physicians (1 = excellent to 4 = non-diagnostic). Of 710 coronary segments assessed on 128-DSCT, 216 (30.4%) achieved an image quality score 1 excellent, 400 (56.3%) score 2, 76 (10.7%) score 3 and 18 (2.6%) score 4 (non-diagnostic). Of 737 coronary segments evaluated on 64-MSCT 271 (36.8%) had an image quality score of 1, 327 (44.4%) 2, 110 (14.9%) score 3, and 29 (3.9%) segments score 4. Average image quality score for both scanners was similar ($P = 0.641$). The mean heart rate during scanning was 58.7 ± 5.6 bpm on 128-DSCT and 59.0 ± 5.6 bpm on 64-MSCT, respectively. Mean effective radiation dose was 1.0 ± 0.2 mSv for 128-DSCT and 1.7 ± 0.6 mSv for 64-MSCT ($P < 0.001$). 128-DSCT with high-pitch spiral mode allows CCTA acquisition with reduced radiation dose at maintained image quality compared to 64-MSCT.

DOI: <https://doi.org/10.1007/s10554-011-9921-3>

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-51125>

Journal Article

Published Version

Originally published at:

Ghadri, J R; Küest, S M; Goetti, R; Fiechter, M; Pazhenkottil, A P; Nkoulou, R N; Kuhn, F P; Pietsch, C; von Schulthess, P; Gaemperli, O; Templin, C; Kaufmann, P A (2012). Image quality and radiation dose comparison of prospectively triggered low-dose CCTA: 128-slice dual-source high-pitch spiral versus 64-slice single-source sequential acquisition. *International Journal of Cardiovascular Imaging*, 28(5):1217-1225.

DOI: <https://doi.org/10.1007/s10554-011-9921-3>

Image quality and radiation dose comparison of prospectively triggered low-dose CCTA: 128-slice dual-source high-pitch spiral versus 64-slice single-source sequential acquisition

Jelena R. Ghadri · Silke M. Küest · Robert Goetti · Michael Fiechter ·
Aju P. Pazhenkottil · Rene N. Nkoulou · Felix P. Kuhn · Carsten Pietsch ·
Patrick von Schulthess · Oliver Gaemperli · Christian Templin · Philipp A. Kaufmann

Received: 7 June 2011 / Accepted: 4 July 2011 / Published online: 10 July 2011
© Springer Science+Business Media, B.V. 2011

Abstract Currently 64-multislice computed tomography (MSCT) scanners are the most widely used devices allowing low radiation dose coronary CT angiography (CCTA) with prospective ECG triggering. Latest 128-slice dual-source CT (DSCT) scanners offer prospective high-pitch spiral acquisition covering the heart during one single beat. We compared radiation dose and image quality from prospective 64-MSCT versus high-pitch spiral 128-slice DSCT scanning, as such data is lacking. CCTA of 50 consecutive patients undergoing 128-DSCT ($2 \times 64 \times 0.6$ mm collimation, 0.28 s rotation time, 3.4 pitch, 100–120 kV tube voltage and 320 mAs tube current–time product) were compared to CCTA of 50 heart rate (HR) and BMI matched patients undergoing 64-MSCT (64×0.625 mm collimation, 0.35 s rotation time, 100–120 kV tube voltage and 400–650 mA

tube current). Image quality was rated on a 4-point scale by two independent cardiac imaging physicians (1 = excellent to 4 = non-diagnostic). Of 710 coronary segments assessed on 128-DSCT, 216 (30.4%) achieved an image quality score 1 excellent, 400 (56.3%) score 2, 76 (10.7%) score 3 and 18 (2.6%) score 4 (non-diagnostic). Of 737 coronary segments evaluated on 64-MSCT 271 (36.8%) had an image quality score of 1, 327 (44.4%) 2, 110 (14.9%) score 3, and 29 (3.9%) segments score 4. Average image quality score for both scanners was similar ($P = 0.641$). The mean heart rate during scanning was 58.7 ± 5.6 bpm on 128-DSCT and 59.0 ± 5.6 bpm on 64-MSCT, respectively. Mean effective radiation dose was 1.0 ± 0.2 mSv for 128-DSCT and 1.7 ± 0.6 mSv for 64-MSCT ($P < 0.001$). 128-DSCT with high-pitch spiral mode allows CCTA acquisition with reduced radiation dose at maintained image quality compared to 64-MSCT.

Jelena R. Ghadri and Silke M. Küest equally contributed to this work.

J. R. Ghadri · S. M. Küest · R. Goetti · M. Fiechter ·
A. P. Pazhenkottil · R. N. Nkoulou · F. P. Kuhn ·
C. Pietsch · P. von Schulthess · O. Gaemperli ·
C. Templin · P. A. Kaufmann (✉)

Department of Radiology, Cardiac Imaging, University
Hospital Zurich, Ramistrasse 100, NUK C 42,
8091 Zurich, Switzerland
e-mail: pak@usz.ch

M. Fiechter · P. A. Kaufmann
Zurich Centre for Integrative Human Physiology (ZIHP),
University of Zurich, Zurich, Switzerland

Keywords Coronary artery disease ·
64-Slice MSCT · 128-Slice DSCT · Image quality ·
Radiation dose

Introduction

The clinical use of coronary computed tomography angiography (CCTA) in patients with low to intermediate risk for coronary artery disease (CAD) has substantially increased over the past decade. This is

not only due to the widespread availability and accessibility of this technique, but it has been substantially facilitated by the introduction of prospective ECG-triggering. This has resulted in a massive radiation dose reduction from over 30 mSv [1] to around 2 mSv and therefore referred to as low-dose CT [2–4] by confining the CT-scan to the smallest possible window at only one distinct mid-diastolic phase of the heart cycle while the X-ray tube is turned off during the rest of the cycle. The beauty of this prospectively gated sequential protocol originally reported for 64-slice CT (MSCT) [2] is the fact that it can be applied to different types of scanners from all vendors including latest generation technology such as 128- or 320-slice CT [5–7].

Recently, the prospectively triggered high-pitch spiral mode has been described as an alternative acquisition protocol for low-dose CCTA [8–10]. However, in order to scan the entire heart in a spiral mode within one single cardiac cycle, a very high-pitch of >3 is necessary even when using a dual-source CT (DSCT) scanner equipped with two 128-slice acquisition detectors. In patients with very low heart rates this highly sophisticated technique can reduce radiation dose for CCTA to 1 mSv or below which is lower than the values reported for sequential scanning with the most widely used 64-slice scanner generation. However, as most centres are not equipped with CT devices from different technology generations and across different vendors, direct comparison is lacking.

Therefore, the aim of the present study was to prospectively compare CCTA image quality and radiation dose from 128-DSCT with prospective high-pitch spiral mode versus 64-MSCT with prospectively triggered sequential scanning.

Materials and methods

Study population

We included 50 consecutive patients undergoing CCTA by 128-DSCT scanning and 50 patients matched for heart rate and BMI who were scanned with 64-MSCT to exclude possible CAD or to evaluate known or suspected CAD. Patients were included if they had signed informed consent authorizing their records to be included in our CCTA research registry.

Exclusion criteria were: allergy to contrast agent, nephropathy (creatinine level >150 $\mu\text{mol/l}$, 1.7 mg/dl), nonsinus rhythm, heart rate >70 bpm after intravenous beta-blocker administration, hemodynamic instability or pregnancy.

Before the study, a detailed interview was conducted to collect data on symptoms, previous cardiac events, and cardiovascular risk factors and collected in our CCTA registry.

CT acquisition and image reconstruction

Before CT scanning all patients received a single sublingual dose of isosorbide dinitrate 2.5 mg (Isoket; Schwarz Pharma, Monheim, Germany) 2 min before scanning [4]. Intravenous metoprolol (Beloc, AstraZeneca, London, UK) was administered to achieve a heart rate less than 70 bpm, as previously reported [7].

CCTA scanning was performed on a second-generation DSCT acquiring 128-slices by use of a z-flying focal spot (128-DSCT, Somatom Definition Flash, Siemens Healthcare, Forchheim, Germany) and a 64-slice MSCT (LightSpeed VCT XT, GE Healthcare, Milwaukee, USA). All patients were instructed about breath holding and the importance of immobility during scanning. All scans on both scanners were performed in cranio-caudal direction with prospective electrocardiogram ECG-triggering. Scanning parameters for 128-DSCT were as follows: center of data acquisition starting at 60% of the RR interval, slice collimation $2 \times 64 \times 0.6$ mm by means of a z-flying focal spot, gantry rotation time 280 ms, pitch 3.4, tube voltage 100 or 120 kV ($\text{BMI} < 25 \text{ kg/m}^2$ or $\text{BMI} > 25 \text{ kg/m}^2$), tube current–time product 320 mAs. A fixed intravenous bolus of 80 ml iopromide (Ultravist 300, 370 mg/ml, Bayer Schering Pharma, Berlin, Germany) was administered at a flow rate of 5 ml/s followed by 60 ml saline solution [11]. Contrast injection was timed by bolus tracking in a sampling region of interest the ascending aorta with the scan beginning 10 s after exceeding a threshold of 100 Hounsfield units (HU). Images were reconstructed with a slice thickness of 0.8 mm and increment of 0.4 mm using a medium smooth kernel designed for cardiac imaging (B26f). All images were anonymised and transferred to an external workstation (AW 4.4, GE Healthcare) for evaluation.

For 64-MSCT a sequential scanning mode with the center of acquisition at 75% of the RR-interval without padding was applied with the following scanning parameters [12]: slice collimation 64×0.625 mm; z-coverage, 40 mm with an increment of 35 mm, gantry rotation time, 350 ms; tube voltage and tube current adapted to BMI (26.3 ± 3.7). Images were reconstructed with a slice thickness of 0.625 mm, using a medium-soft tissue convolution kernel (standard). We used a BMI-adapted contrast material protocol, which has been previously established [10, 13, 14].

Of note, according to our standard routine protocol, no additional coronary calcium scanning was performed.

Image quality of coronary artery segments

According to the 16-segment model of the American Heart Association [15]. All coronary segments were assigned to one of the main coronary arteries as follows: right coronary artery (RCA) segments 1–4, left main artery (LMA) and the left anterior ascending artery (LAD) segments 5–10, and the left circumflex artery (LCX) segments 11–15, intermediate artery segment 16, if present. The image quality was visually assessed by two independent readers for each segment on a 4-point scale using axial source images and multiplanar reformations, whereby score 1 = excellent (no motion artefacts); score 2 = good (minor artefacts); score 3 = adequate (moderate artefacts); score 4 = non-diagnostic, as previously reported [3]. For any disagreement in data evaluation between the two readers, consensus agreement was achieved.

Image noise and contrast-to-noise ratio (CNR) were determined by one observer not involved in the image quality score rating. The vessel contrast was measured as the difference in mean attenuation in HU in the contrast enhanced vessel lumen and the mean attenuation in HU in the adjacent perivascular tissue. Attenuation was measured in round-shaped regions of interest (ROI) in the proximal LMA and RCA (measurements 1 and 2). Image noise was defined as the SD of the attenuation value in a ROI placed in the ascending aorta. All ROIs were drawn as large as possible, carefully avoiding the vessel wall and blooming halo of calcifications. CNR was calculated as the ratio of the mean of measurements 1 and 2, and image noise as previously described [16, 17].

Radiation dose

Values for effective radiation dose were calculated by multiplying the dose length product (DLP) with a conversion factor for the chest ($k = 0.014$ mSv/mGyxcn) as previously suggested [18] and adopted in large trials[1].

Statistical analysis

Statistical analysis was performed using the SPSS software (19.0, SPSS inc., Chicago IL, USA). Quantitative variables were expressed as mean \pm standard deviation (SD) or median (range) as appropriate, and categorical variables as frequencies, or percentages.

The Mann–Whitney U test was performed to analyse the differences between the two groups regarding image quality, radiation dose, image noise and contrast-to-noise ratio. The student's *t* test or Mann–Whitney U test was used to compare the means of the continuous variables (demographic data) and contingency tables (risk factors, referral indications, CAD data. Image artefact per patient) were analyzed using chi-square test.

Interobserver agreement was evaluated by Cohen's Kappa coefficient. A *k* value of greater of 0.81 was defined as excellent, 0.61–0.80 was considered as good, values of 0.41–0.60 as moderate and values below 0.20 as poor agreement.

A two-tailed *P* value < 0.05 was considered statistically significant. Comparison of patient characteristics was evaluated with Chi-squared test for unnormal distribution. The impact on mean heart rate on image quality was assessed by Spearman's rank-order. Correlation coefficients were calculated to compare HR with the mean image quality scores of all coronary segments on a per patient basis.

Results

Study population

All 100 patients underwent successful scanning with either high-pitch acquisition on 128-DSCT or sequential scanning on 64-MSCT. The mean age of the 128-DSCT group was 57.5 ± 13.1 years and 59.9 ± 12.2 for the 64-MSCT group ($P = 0.46$). Patients were adequately matched regarding heart rate ($P = 0.69$)

and BMI ($P = 0.81$) and there were no significant differences between the two groups regarding cardiovascular risk factors. The patient baseline characteristics, referral indications and previous state of CAD are given in Table 1.

Subjective image quality

Image quality was assessed for all 1,447 coronary segments from 128-DSCT ($n = 710$) and for 64-MSCT ($n = 737$) (Table 2). The total number of segments is not equal due to anatomic variations with not all segments being present in all patients. Interobserver agreement for rating image quality

was good for high-pitch spiral mode (128-DSCT; $k = 0.66$) and for sequential scanning (64-MSCT; $k = 0.61$).

Among the 128-DSCT scans 692 (97.5%) segments yielded diagnostic image quality (score 1–3), which was excellent in 216 (30.4%), good in 400 (56.3%) and adequate in 76 (10.7%) coronary segments (Table 3). Similarly among the 64-MSCT scans 708 (96.1%) coronary segments revealed diagnostic image quality (score 1–3), which was excellent in 271 (36.8%), good in 327 (44.4%) and adequate in 110 (14.9%) coronary segments (Table 3). Figure 1 demonstrates the frequency of image quality for overall coronary segments for 128-DSCT and 64-MSCT.

Causes for poor image quality were the presence of motion or step-artefacts. None of the patients scanned with 128-DSCT revealed step-artefacts, however motion artefacts occurred in 14 patients

Table 1 Baseline characteristics

Demographics	128-DSCT <i>n</i> = 50	64-MSCT <i>n</i> = 50	<i>P</i>
Male	35 (70%)	33 (66%)	0.409
Age (years)	57.5 ± 13.1	59.9 ± 12.2	0.469
Weight (kg)	77.4 ± 11.9	78.6 ± 14.5	0.735
Height (cm)	171.4 ± 9.5	172.5 ± 9.1	0.756
BMI (kg/m ²)	26.4 ± 3.4	26.3 ± 3.7	0.817
Heart rate during scan (bpm)	58.7 ± 5.6	59.0 ± 5.6	0.694
<i>Risk factors</i>			
Hypertension	27 (54%)	31 (62%)	0.417
Dyslipidaemia	25 (50%)	24 (48%)	0.841
Diabetes mellitus	4 (8%)	5 (10%)	0.726
Smoking	14 (28%)	22 (44%)	0.095
Family history	19 (38%)	19 (38%)	1.000
<i>Referral indication</i>			
Chest pain	11 (22%)	10 (20%)	0.814
Atypical chest pain	5 (10%)	17 (34%)	0.004
None	11 (22%)	11 (22%)	1.000
Dyspnoea	8 (16%)	2 (4%)	0.045
Pre-operative assessment	8 (16%)	4 (8%)	0.218
Bicycle test pathologic	6 (12%)	4 (8%)	0.444
Check-up	1 (2%)	2 (4%)	0.557
<i>Coronary artery disease (CAD)</i>			
Known CAD	7 (14%)	11 (22%)	0.297
Previous bypass	4 (8%)	4 (8%)	1.000
Previous stenting	3 (6%)	5 (10%)	0.461
Prior myocardial infarction	1 (2%)	6 (12%)	0.050

BMI body mass index, CAD coronary artery disease


Table 2 Mean image quality score

	128-DSCT	64-MSCT
<i>Per segment</i>		
Segment 1	1.65	1.60
Segment 2	2.12	2.44
Segment 3	2.58	2.57
Segment 4	1.90	1.71
Segment 5	1.37	1.24
Segment 6	1.40	1.38
Segment 7	1.76	1.86
Segment 8	2.18	2.32
Segment 9	1.90	1.78
Segment 10	1.93	2.00
Segment 11	1.62	1.35
Segment 12	1.74	1.77
Segment 13	1.74	1.72
Segment 14	2.00	2.09
Segment 15	2.00	2.44
Segment 16	1.87	1.67
<i>Per vessel</i>		
RCA	2.06	2.09
LMA/LAD	1.75	1.76
LCX	1.80	1.80
<i>Per patient</i>	1.85	1.86

RCA right coronary artery, LMA left main artery, LAD left anterior descending artery, LCX left circumflex artery

Table 3 Image Quality total (128-DSCT $n = 710$; 64-MSCT $n = 737$; total segments $n = 1447$)

Total segments	128-DSCT 710	64-MSCT 737	<i>P</i>
Score 1	216 (30.4%)	271 (36.8%)	0.299
Score 2	400 (56.3%)	327 (44.4%)	0.03
Score 3	76 (10.7%)	110 (14.9%)	0.256
Score 4	18 (2.6%)	29 (3.9%)	0.984

**Fig. 1** Frequency of image quality for overall coronary segments for 128-DSCT (white bars) and 64-MSCT (black bars), (* $P = 0.03$; ns not significant)**Table 4** Image artifact per patient


	128-DSCT $n = 50$	64-MSCT $n = 50$	<i>P</i>
Step artefact	None	9 (18%)	0.001
Motion artefact	14 (28%)	10 (20%)	0.349

(28%). In the 64-MSCT group nine patients (18%) revealed step artefacts, and 10 (20%) motion artefacts (Table 4).

There was no significant difference ($P = 0.641$) in image quality between the two scanners with different scanning protocols for any of the coronary segments as well as for the overall image quality.

Impact of heart rate on image quality

The average heart during scanning was 58.7 ± 5.6 bpm (range 46–70 bpm) for 128-DSCT and 59.0 ± 5.6 bpm (range 47–70 bpm) for 64-MSCT.

**Fig. 2** Linear regression plot of mean image quality scores over all coronary segments per patient (y-axis) against heart rate during CT-scanning (x-axis) in 128-DSCT **a** and 64-MSCT **b**. The two outer lines represent 95% individual prediction interval (Spearman's correlation for 128-DSCT $r = 0.11$, $P = 0.45$; Spearman's correlation for 64-MSCT $r = 0.68$, $P < 0.01$)

There was no significant correlation between HR and mean image quality score on per patient analysis ($r = 0.11$, $P = 0.45$) for 128-DSCT, while a significant correlation between heart rate and image quality was found for 64-MSCT scanning ($r = 0.68$, $P \leq 0.01$), Fig. 2a and b.

Table 5 Radiation dose

	128-DSCT	64-MSCT	<i>P</i>
CTDI (mGy)	3.3 ± 0.8	8.2 ± 2.8	<0.001
DLP (mGyxcn)	71.2 ± 17.7	124.5 ± 41.7	<0.001
Effective radiation dose (mSv)	1.0 ± 0.3	1.7 ± 0.6	<0.001
Scan length (cm)	14.4 ± 2.6	14.9 ± 2.2	0.306

CTDI Computed tomography dose index, *DLP* dose length product

Objective image quality

Contrast was significantly different between 128-DSCT and 64 MSCT (532.7 ± 122.0 vs. 502.8 ± 84.0 ; $P = 0.014$). Image noise was not significantly different (128-DSCT: 31.5 ± 8.3 HU; 64-MSCT: 34.6 ± 8.8 HU; $p = 0.056$). CNR was (18.3 ± 5.5 vs. 14.6 ± 4.4 ; $P < 0.01$).

Radiation dose

There was a significant difference in the effective radiation dose between the 128-DSCT high-pitch group (1.0 ± 0.3 mSv) and the 64-MSCT sequential scanning group (1.7 ± 0.6 mSv) ($P > 0.001$). The average DLP was 71.2 ± 17.7 (mGyxcn) on 128-DSCT and 124.5 ± 41.7 (mGyxcn) on 64-MSCT, although the scan length did not differ significantly between the two (14.4 ± 2.6 cm vs. 14.9 ± 2.2 cm, $P = \text{n.s.}$) (Table 5). The mean tube current–time product/tube voltage was 383 ± 43 mAs/107 ± 10 kV for 64-MSCT and 327 ± 22 mAs/101 ± 5 kV for 128-DSCT.

Discussion

The recent years have witnessed tremendous technological advances in multislice CT systems enabling high image quality CCTA in clinical practice while constantly reducing radiation exposure for the patient. Since there is no threshold below which radiation exposure can be proven to not be potentially cancerogenous, any radiation dose can be potentially harmful and should therefore be minimized [19].


This is the first study with a cross-vendor comparison of CCTA image quality and radiation dose from 128-DSCT versus 64-MSCT. In fact, much effort has

been invested in developing methods to improve image quality and/or reduce radiation dose in CCTA. Unfortunately, most approaches to solve one issue act in opposite direction on the other issue. Our results document that CCTA with 128-DSCT using high-pitch scanning allows further substantial reduction in radiation exposure compared to low-dose CCTA with prospective sequential scanning. This was achieved at maintained image quality, which is of great importance in the context mentioned above (Fig. 3). The amount of radiation dose reduction is in line with previous reports of radiation dose saving with new generation scanners [7, 20]. Recent data have demonstrated the feasibility of prospective ECG-triggering with high-pitch spiral CCTA acquisition with sub-mSv radiation doses [8]. However, none of these studies have compared radiation doses and image quality of CCTA from different systems of different vendors.

In our study we found a difference of the mean effective radiation dose between the two groups of 0.7 mSv, which implies a dose reduction of 41.2% when using the high-pitch 128-DSCT spiral protocol. Although an absolute decrease in radiation dose of 0.7 mSv may appear rather modest with regards to the massive increase in cost from single to dual source scanners, even this radiation dose saving may be worth its cost, as the radiation dose for a CCTA now is as low as a fraction of the annual background exposure of approximately 3.2 mSv by radon [21]. In fact, since there is no threshold below which radiation has been proven to not be cancerogenous, any amount of radiation must be considered potentially harmful. Consequently, any amount of dose reduction should be considered worth the cost.

The observed dose reduction despite the same scan length (which otherwise is a major determinant of the dose-length-product, DLP), may be explained considering various factors. First, tube voltage and tube current settings are not directly comparable between the two protocols used in this study. However, we have used well established protocols specific for high-pitch spiral [7, 20] and prospectively triggered sequential scanning [22]. Second, inherent differences between spiral and sequential protocol may have contributed to the dose reduction. For example oversampling is minimized using the high-pitch dual-spiral technique, while a certain overlap of the individual blockwise scans that comprise a study is necessary for precise image reconstruction in

Fig. 3 Prospectively ECG-triggered images on 128-DSCT (a–c) and 64-MSCT (d–f) of the right coronary artery (RCA) in volume rendered multiplanar reconstruction and curved multiplanar reformation. a–c, shows RCA in a 45 year old patient with an average heart rate of 59/min and a BMI 29.07 kg/m². d–f, shows RCA in a 60 year old patient with a heart rate 59/min and a BMI 29.07 kg/m², respectively. Image quality was scored with 1 (=excellent) for both scanners



64-sequential scanning with 64-MSCT. In addition X-rays irradiating the entire field of view after turning on the beam before starting the scan, expose the patient only once at the begin and at the end of the high-pitch spiral scan. By contrast, with prospective sequential scanning the unnecessary exposure occurs repeatedly for every sequential block.

Studies comparing radiation doses with 128-DSCT have reported even higher dose reduction rates than in the present study, however those comparisons were done against retrospective gating protocols and not versus prospective triggering mode [23]. The image quality which has also been assessed in the study by Duarte et al. [23] evaluated with different scanners 128-DSCT versus 64-MSCT using different scanning modes (high-pitch vs. retrospective scanning) has demonstrated superiority of the prospectively triggered high-pitch mode. This may be at, least in part, due to the fact that sequential scanning can induce stair-step artefacts due to irregular heart beat and can limit the accuracy of CCTA, this does not occur in spiral mode. In fact, in the present study stair step artefacts were observed in 18% of 64-MSCT scans, but not in 128-DSCT scans.

Interestingly, while we found a significant correlation ($r = 0.68$, $P \leq 0.01$) between heart rate and

image quality for 64-MSCT, in line with previous results [3, 24], no such correlation ($r = 0.11$, $P = 0.45$) was found for 128-DSCT. This is probably due to the fact that once heart rate is as low as to ensure a diastolic phase long enough to allow coverage of the whole coronary tree within one heart beat according to the prospective high-pitch spiral protocol, the absolute heart rate has no influence on image quality. By contrast, the prospectively triggered sequential scanning protocol involves several heart beats and requiring very low heart rate variability to avoid stair step artefacts [3, 24]. As increasing heart rates are more likely to be associated with heart rate variability image quality may be affected by higher heart rates explaining the correlation of heart rate and image quality score in 64-MSCT. However, higher heart rates over 70 bpm may have more deleterious impact on image quality in 128-DSCT from 64-MSCT, but heart rates beyond 70 bpm were excluded in the present study.

Study limitations

The image quality was visually assessed by a 4-point-scale and not by an automated system; this however, has been validated and documented to produce

reliable results [3]. The present study does not represent a head-to-head comparison, as two different patient groups had to be identified because repeat scanning of the same patients would not be appropriate due to radiation exposure and ethical issues. However, we took due care to ascertain an optimal match of the two study groups with regard to the parameters identified as most relevant determinants of image quality [3, 25] in CCTA. All images were post-processed on the same workstation/software from one vendor to minimize the potential bias of different workstations/software on image quality, as it has been recently shown that differences in reconstruction algorithm may introduce more variability than different scanners. Finally, slice thickness was not equal in 128-DSCT versus 64-MSCT, which may affect image noise and consequently image quality. However, our study did not reveal any difference in noise from 128-DSCT versus 64-MSCT, excluding a relevant bias due to this technical issue.

Conclusion

128-DSCT with high-pitch spiral mode allows CCTA acquisition with reduced radiation dose at maintained image quality compared to 64-MSCT.

Acknowledgments We are grateful to Ennio Mueller, Edlira Loga, Mirjam De Bloeme, Sabrina Epp, and Daisy Beutel for their excellent technical support. The study was supported by a grant of the Swiss National Science Foundation (SNSF) and by the ZHIP (Zurich Center for Integrative Human Physiology, University of Zurich, Switzerland).

Conflict of interest None.

References

- Hausleiter J, Meyer T, Hermann F et al (2009) Estimated radiation dose associated with cardiac CT angiography. *JAMA* 301:500–507
- Husmann L, Valenta I, Gaemperli O et al (2008) Feasibility of low-dose coronary CT angiography: first experience with prospective ECG-gating. *Eur Heart J* 29:191–197
- Herzog BA, Husmann L, Burkhard N et al (2009) Low-dose CT coronary angiography using prospective ECG-triggering: impact of mean heart rate and heart rate variability on image quality. *Acad Radiol* 16:15–21
- Buechel RR, Husmann L, Herzog BA et al (2011) Low-dose computed tomography coronary angiography with prospective electrocardiogram triggering: feasibility in a large population. *J Am Coll Cardiol* 57:332–336
- Kaufmann PA (2008) Low-dose computed tomography coronary angiography with prospective triggering: a promise for the future. *J Am Coll Cardiol* 52:1456–1457
- Dewey M, Zimmermann E, Deissenrieder F et al (2009) Noninvasive coronary angiography by 320-row computed tomography with lower radiation exposure and maintained diagnostic accuracy: comparison of results with cardiac catheterization in a head-to-head pilot investigation. *Circulation* 120:867–875
- Alkadhi H, Stolzmann P, Desbiolles L et al (2010) Low-dose, 128-slice, dual-source CT coronary angiography: accuracy and radiation dose of the high-pitch and the step-and-shoot mode. *Heart* 96:933–938
- Leschka S, Stolzmann P, Desbiolles L et al (2009) Diagnostic accuracy of high-pitch dual-source CT for the assessment of coronary stenoses: first experience. *Eur Radiol* 19:2896–2903
- Hausleiter J, Bischoff B, Hein F et al (2009) Feasibility of dual-source cardiac CT angiography with high-pitch scan protocols. *J Cardiovasc Comput Tomogr* 3:236–242
- Herzog BA, Husmann L, Burkhard N et al (2008) Accuracy of low-dose computed tomography coronary angiography using prospective electrocardiogram-triggering: first clinical experience. *Eur Heart J* 29:3037–3042
- Goetti R, Feuchtner G, Stolzmann P et al (2010) High-pitch dual-source CT coronary angiography: systolic data acquisition at high heart rates. *Eur Radiol* 20:2565–2571
- Herzog BA, Wyss CA, Husmann L et al (2009) First head-to-head comparison of effective radiation dose from low-dose 64-slice CT with prospective ECG-triggering versus invasive coronary angiography. *Heart* 95:1656–1661
- Pazhenkottil AP, Husmann L, Buechel RR et al (2010) Validation of a new contrast material protocol adapted to body surface area for optimized low-dose CT coronary angiography with prospective ECG-triggering. *Int J Cardiovasc Imag* 26:591–597
- Husmann L, Herzog BA, Burkhard N et al (2009) Low-dose coronary CT angiography with prospective ECG triggering: validation of a contrast material protocol adapted to body mass index. *Am J Roentgenol* 193:802–806
- Austen WG, Edwards JE, Frye RL et al (1975) A reporting system on patients evaluated for coronary artery disease. Report of the Ad Hoc committee for grading of coronary artery disease, council on cardiovascular surgery, American heart association. *Circulation* 51:5–40
- Goetti R, Leschka S, Baumuller S et al (2010) Low dose high-pitch spiral acquisition 128-slice dual-source computed tomography for the evaluation of coronary artery bypass graft patency. *Investig Radiol* 45:324–330
- Husmann L, Herzog BA, Burger IA et al (2010) Usefulness of additional coronary calcium scoring in low-dose CT coronary angiography with prospective ECG-triggering impact on total effective radiation dose and diagnostic accuracy. *Acad Radiol* 17:201–206
- Einstein AJ, Moser KW, Thompson RC et al (2007) Radiation dose to patients from cardiac diagnostic imaging. *Circulation* 116:1290–1305

19. Einstein AJ (2009) Radiation protection of patients undergoing cardiac computed tomographic angiography. *JAMA* 301:545–547
20. Achenbach S, Marwan M, Ropers D et al (2010) Coronary computed tomography angiography with a consistent dose below 1 mSv using prospectively electrocardiogram-triggered high-pitch spiral acquisition. *Eur Heart J* 31:340–346
21. Kaufmann PA, Knuuti J (2011) Ionizing radiation risks of cardiac imaging: estimates of the immeasurable. *Eur Heart J* 32:269–271
22. Husmann L, Herzog BA, Gaemperli O et al (2009) Diagnostic accuracy of computed tomography coronary angiography and evaluation of stress-only single-photon emission computed tomography/computed tomography hybrid imaging: comparison of prospective electrocardiogram-triggering versus retrospective gating. *Eur Heart J* 30:600–607
23. Duarte R, Bettencourt N, Costa JC et al (2010) Coronary computed tomography angiography in a single cardiac cycle with a mean radiation dose of approximately 1 mSv: initial experience. *Rev Port Cardiol* 29:1667–1676
24. Husmann L, Herzog BA, Burkhard N et al (2009) Body physique and heart rate variability determine the occurrence of stair-step artefacts in 64-slice CT coronary angiography with prospective ECG-triggering. *Eur Radiol* 19:1698–1703
25. Tatsugami F, Husmann L, Herzog BA et al (2009) Evaluation of a body mass index-adapted protocol for low-dose 64-MDCT coronary angiography with prospective ECG triggering. *AJR Am J Roentgenol* 192:635–638